



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/781,030	02/17/2004	Rainer Schneider	5497-67414	2830

24197 7590 04/10/2007  
KLARQUIST SPARKMAN, LLP  
121 SW SALMON STREET  
SUITE 1600  
PORTLAND, OR 97204

EXAMINER
----------

LEWIS, RALPH A

ART UNIT	PAPER NUMBER
----------	--------------

3732

SHORTENED STATUTORY PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE
3 MONTHS	04/10/2007	PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

ED

**Office Action Summary**

Application No.

10/781,030

Applicant(s)

SCHNEIDER, RAINER

Examiner

Ralph A. Lewis

Art Unit

3732

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 17 January 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1-39 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-39 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
 Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
 Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- |  |   |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)                     | 4) <input type="checkbox"/> Interview Summary (PTO-413)           |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____                                      |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)          | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____  | 6) <input type="checkbox"/> Other: _____                          |

### **Rejections based on Prior Art**

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-29, 31-33, 35-37 and 39 are rejected under 35 U.S.C. 102(b) as being anticipated by Bianchetti et al (EP 1103232).

Bianchetti et al disclose in Figure 4 an adapter body 101 having electrical leads 106 that is positioned between a supply conduit 8 and dental light hand piece 103. Electrical power is supplied to the adapter 101 from the supply conduit 8 where it in turn is supplied to the hand piece 103 by adapter electrodes (leads) 106 disposed diametrically opposite each other (only one visible in Figure 4, column 6, lines 54-57). The hand piece 103 includes first and second electrical contacts 111 and 112 that connect with the adapter electrodes 106 when connected with the hand piece 103. The adapter 101 is symmetrical with electrodes 106 on opposite sides from one another, consequently it is capable of being rotated manually 180 degrees and inserted into hand piece 103 thereby switching the polarity of the adapter electrodes 106 in contact with the handpiece. Moreover, it is noted that the hand piece 103 includes LEDS (column 7, lines 3-7) which inherently and necessarily requires the proper polarity in order to operate.

In regard to claim 5, the Bianchetti et al electrode leads 106 slide into contact with the handpiece contacts 111 and 112 when being connected.

In regard to claim 10, the Bianchetti et al connection acts as a basic switch.

In response to the present rejection applicant amended the claims to require that the adapter leads were selectively switchable by a manual operation relative to the supply conduit. The examiner notes that one could manually disassemble the adapter 101 and manually switch the electrical leads. Applicant's "switchable by a manual operation" language is quite broad.

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 30, 34 and 38 are rejected under 35 U.S.C. 103(a) as being unpatentable over Bianchetti et al (EP 1103232).

Bianchetti et al do not explicitly disclose what happens if a user connects the hand piece 103 to the adapter 101 with the electrical connections inadvertently reversed. One of ordinary skill in the art, however, would have readily understood that when the LEDs are not hooked up with the proper polarity then they will not operate. It naturally follows, then that if the Bianchetti et al device were inadvertently hooked up with the wrong polarities through adapter 101, that one would rotate the adapter 101

Art Unit: 3732

with respect to the hand piece 103 to reverse the polarities so that the device would then function.

#### **Action Made Final**

**THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the mailing date of this final action.


Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should

Art Unit: 3732

you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Any inquiry concerning this communication should be directed to **Ralph Lewis** at telephone number **(571) 272-4712**. Fax (571) 273-8300. The examiner works a compressed work schedule and is unavailable every other Friday. The examiner's supervisor, Cris Rodriguez, can be reached at (571) 272-4964.

R. Lewis  
April 2, 2007



Ralph A. Lewis  
Primary Examiner  
Au 3732